

Product specifications

To learn more about our contact lenses visit:
www.coopervision.com



Product	Sphere Power	Cylinder Power (DC)	Axis°	Add Power (D)	Design	Technologies	Wear Schedule & Replacement Frequency	Base Curve (mm)	Diameter (mm)	Material USAN	Water content (%)	Oxygen transmissibility Dk/t†	UV Blocking* & Class	Visibility Tint	FDA Group
 clariti® 1 day sphere	-0.50D to -10.00D +0.50D to +8.00D (0.50D steps after +/- 6.00D) No Plano	— — —	— — —	— — —	Asphere	WetLoc® Technology	Daily wear; 1 day replacement	8.6	14.1	somofilcon A	56	86	Class 2	No	5B (SiHy)
 clariti® 1 day multifocal	+5.00D to -6.00D (0.25D steps)	— — —	— — —	Low: up to +2.25D High: +2.50D to +3.00D	Multifocal	WetLoc® Technology	Daily wear; 1 day replacement	8.6	14.1	somofilcon A	56	86	Class 2	No	5B (SiHy)
 clariti® 1 day multifocal 3 Add	+8.00D to -10.00D (0.25D steps) -10.50D to -12.00D (0.50D steps)	— — —	— — —	Low: +0.75D to +1.25D Med: +1.50D to +1.75D High: +2.00D to +2.50D	Multifocal	WetLoc® Technology Binocular Progressive System™	Daily wear; 1 day replacement	8.6	14.1	somofilcon A	56	86	Class 2	No	5B (SiHy)
 clariti® 1 day toric	Plano to -6.00D (0.25D steps)	-0.75, -1.25, -1.75	Full circle in 10° steps	— — —	Toric	Prism ballast design WetLoc® Technology	Daily wear; 1 day replacement	8.6	14.3	somofilcon A	56	57	Class 2	No	5B (SiHy)
		-2.25	10°, 20°, 70°, 80°, 90°, 100°, 110°, 160°, 170°, 180°												
	-6.50D to -9.00D (0.50D steps)	-0.75, -1.25, -1.75	10°, 20°, 60°, 70°, 80°, 90°, 100°, 110°, 120°, 160°, 170°, 180°												
	+0.25D to +4.00D (0.25D steps)	-0.75, -1.25, -1.75	10°, 20°, 90°, 160°, 170°, 180°												

† (@-3.00DS) x 10⁻⁹ [(cm/sec) x (ml O₂)/(ml x mm Hg)]

* UV-blocking contact lenses help provide protection against transmission of harmful UV radiation to eye but are not substitutes for protective UV-absorbing eyewear, such as UV-absorbing goggles or sunglasses, as they do not completely cover the eye or surrounding area. Continue to use UV-absorbing eyewear as directed by your eye care professional.

Product specifications

To learn more about our contact lenses visit:
www.coopervision.com



Product	Sphere Power	Cylinder Power (DC)	Axis°	Add Power (D)	Design	Technologies	Wear Schedule & Replacement Frequency	Base Curve (mm)	Diameter (mm)	Material USAN	Water content (%)	Oxygen transmissibility Dk/t†	UV Blocking* & Class	Visibility Tint	FDA Group
 MiSight® 1 day	-0.50D to -7.00D (-0.25D steps, -0.50D after -6.00D) No plano	— —	— —	Treatment zones offer +2.00D of myopic defocus	Dual-focus	ActivControl® Technology PC Technology™ (phosphoryl choline)	Daily wear; 1 day replacement	8.7	14.2	omafilcon A	60	28	No	Yes	2
 MyDay® daily disposable	+8.00D to -12.00D (0.50D steps after +5.00D and -6.00D) No Plano	— —	— —	— —	Asphere	Aberration Neutralizing System™ Aquaform® Technology	Daily wear; 1 day replacement	8.4	14.2	stenfilcon A	54	100	Class 2	Yes	5B (SiHy)
 MyDay Energys®	+8.00D to -12.00D (0.50D steps after +5.00D and -6.00D) No Plano	— —	— —	— —	Asphere	DigitalBoost™ Aquaform® Technology	Daily wear; 1 day replacement	8.4	14.2	stenfilcon A	54	100	Class 2	Yes	5B (SiHy)
 MyDay® multifocal	+8.00D to -10.00D (0.25D steps) -10.50D to -12.00D (0.50D steps)	— —	— —	Low: +0.75D to +1.25D Med: +1.50D to +1.75D High: +2.00D to +2.50D	Multifocal	Binocular Progressive System® Aquaform® Technology	Daily wear; 1 day replacement	8.4	14.2	stenfilcon A	54	100	Class 2	Yes	5B (SiHy)
 MyDay® toric	+6.00D to -10.00D (0.50D steps after -6.00D)	-0.75, -1.25, -1.75, -2.25	Full circle in 10° steps	— —	Toric	Optimized Toric Lens Geometry™ Aquaform® Technology	Daily wear; 1 day replacement	8.6	14.5	stenfilcon A	54	80	Class 2	Yes	5B (SiHy)

† (@-3.00DS) x 10⁻⁹ [(cm/sec) x (ml O₂)/(ml x mm Hg)]

* UV-blocking contact lenses help provide protection against transmission of harmful UV radiation to eye but are not substitutes for protective UV-absorbing eyewear, such as UV-absorbing goggles or sunglasses, as they do not completely cover the eye or surrounding area. Continue to use UV-absorbing eyewear as directed by your eye care professional.
 Biomedics®, clariti®, CooperVision®, MyDay®, MyDay Energys®, live®, MiSight® and Proclear® are registered trademarks of The Cooper Companies, Inc. and its subsidiaries.

Product specifications

To learn more about our contact lenses visit:
www.coopervision.com



Product	Sphere Power	Cylinder Power (DC)	Axis°	Add Power (D)	Design	Technologies	Wear Schedule & Replacement Frequency	Base Curve (mm)	Diameter (mm)	Material USAN	Water content (%)	Oxygen transmissibility Dk/t†	UV Blocking* & Class	Visibility Tint	FDA Group
 Proclear® 1 day	+8.00D to -12.00D (0.50D steps after +5.00D and -6.00D) No Plano	— — —	— — —	— — —	Asphere	Aberration Neutralizing System™ PC Technology™	Daily wear; 1 day replacement	8.7	14.2	omafilcon A	60	28	No	Yes	2
 Proclear® 1 day multifocal	+6.00D to -10.00D (0.50D steps after -6.00D)	— — —	— — —	Single power profile	Multifocal	PC Technology™	Daily wear; 1 day replacement	8.7	14.2	omafilcon A	60	28	No	Yes	2

*† (@-3.00DS) $\times 10^{-9}$ [(cm/sec) $\times$ (ml O₂)/(ml $\times$ mm Hg)]

† UV-blocking contact lenses help provide protection against transmission of harmful UV radiation to eye but are not substitutes for protective UV-absorbing eyewear, such as UV-absorbing goggles or sunglasses, as they do not completely cover the eye or surrounding area. Continue to use UV-absorbing eyewear as directed by your eye care professional.
 Biomedics®, clariti®, CooperVision®, MyDay®, MyDay Energys®, live®, MiSight® and Proclear® are registered trademarks of The Cooper Companies, Inc. and its subsidiaries.

Product specifications

To learn more about our contact lenses visit:
www.coopervision.com



Product	Sphere Power	Cylinder Power (DC)	Axis°	Add Power (D)	Design	Technologies	Wear Schedule & Replacement Frequency	Base Curve (mm)	Diameter (mm)	Material USAN	Water content (%)	Oxygen transmissibility Dk/t†	UV Blocking* & Class	Visibility Tint	FDA Group
 Avaira Vitality™	+8.00D to -12.00D (0.50D steps after +/-6.00D) No Plano	— — —	— — —	— — —	Asphere	Aberration Neutralizing System™ Aquaform® Technology	Daily wear; 2 weekly replacement	8.4	14.2	fanfilcon A	55	110	Class 1	Yes	5B (SiHy)
 Avaira Vitality™ toric	+8.00D to -10.00D (0.50D steps after ± 6.00D)	-0.75 -1.25 -1.75 -2.25	10° to 180° (10° steps)	— — —	Toric	Optimized Toric Lens Geometry™ Aquaform® Technology	Daily wear; 2 weekly replacement	8.5	14.5	fanfilcon A	55	90	Class 1	Yes	5B (SiHy)
 Biofinity®	+8.00D to -12.00D (0.50D steps after ±6.00D) No Plano	— — —	— — —	— — —	Asphere	Aberration Neutralizing System™ Aquaform® Technology	Daily wear or extended wear 7 days/6 nights; 30 day replacement	8.6	14.0	comfilcon A	48	160	No	Yes	5C (SiHy)
 Biofinity Energys®	+8.00D to -12.00D (0.50D steps after ±6.00D) No Plano	— — —	— — —	— — —	Asphere	DigitalBoost™ Aquaform® Technology	Daily wear or extended wear 7 days/6 nights; 30 day replacement	8.6	14.0	comfilcon A	48	160	No	Yes	5C (SiHy)

† (@-3.00DS) x 10⁻⁹ [(cm/sec) x (ml O₂)/(ml x mm Hg)]

* UV-blocking contact lenses help provide protection against transmission of harmful UV radiation to eye but are not substitutes for protective UV-absorbing eyewear, such as UV-absorbing goggles or sunglasses, as they do not completely cover the eye or surrounding area. Continue to use UV-absorbing eyewear as directed by your eye care professional.

Product specifications

To learn more about our contact lenses visit:
www.coopervision.com



Product	Sphere Power	Cylinder Power (DC)	Axis°	Add Power (D)	Design	Technologies	Wear Schedule & Replacement Frequency	Base Curve (mm)	Diameter (mm)	Material USAN	Water content (%)	Oxygen transmissibility Dk/t†	UV Blocking* & Class	Visibility Tint	FDA Group
 Biofinity® multifocal	+6.00D to -10.00D (0.50D steps after -6.00D)	— —	— —	+1.00 +1.50 +2.00 +2.50	Multifocal D lens N Lens	Balanced Progressive™ Technology Aquaform® Technology	Daily wear or extended wear 7 days/6 nights; 30 day replacement	8.6	14.0	comfilcon A	48	142	No	Yes	5C (SiHy)
 Biofinity® toric	+8.00D to -10.00D (0.50D steps after +/-6.00D)	-0.75 -1.25 -1.75 -2.25	10° to 180° (10° steps)	— —	Toric	Optimized Toric Lens Geometry™ Aquaform® Technology	Daily wear or extended wear 7 days/6 nights; 30 day replacement	8.7	14.5	comfilcon A	48	116	No	Yes	5C (SiHy)
 Biofinity® toric multifocal	+10.00D to -10.00D (0.50D steps after +/-6.00D)	-0.75 to -5.75 (0.50D steps)	5° to 180° (5° steps)	+1.00 +1.50 +2.00 +2.50	Toric Multifocal D Lens N Lens	Optimized Toric Lens Geometry™ Balanced Progressive™ Technology Aquaform® Technology	Daily wear or extended wear 7 days/6 nights; 30 day replacement	8.7	14.5	comfilcon A	48	116	No	Yes	5C (SiHy)
 Biofinity® XR	+8.50D to +15.00D -12.50D to -20.00D (0.50D steps)	— —	— —	— —	Asphere	Aberration Neutralizing System™ Aquaform® Technology	Daily wear or extended wear 7 days/6 nights; 30 day replacement	8.6	14.0	comfilcon A	48	160	No	Yes	5C (SiHy)
 Biofinity® XR Toric	+8.50D to +20.00D and -10.50D to -20.00D (0.50D steps) +20.00D to -20.00D (0.50D steps after +/-6.00D)	-0.75 -1.25 -1.75 -2.25 -2.75 -3.25 -3.75 -4.25 -4.75 -5.25 -5.75	5° to 180° (5° steps)	— —	Toric	Optimized Toric Lens Geometry™ Aquaform® Technology	Daily wear or extended wear 7 days/6 nights; 30 day replacement	8.7	14.5	comfilcon A	48	116	No	Yes	5C (SiHy)

† (@-3.00DS) x 10⁻⁹ [(cm/sec) x (ml O₂)/(ml x mm Hg)]

* UV-blocking contact lenses help provide protection against transmission of harmful UV radiation to eye but are not substitutes for protective UV-absorbing eyewear, such as UV-absorbing goggles or sunglasses, as they do not completely cover the eye or surrounding area. Continue to use UV-absorbing eyewear as directed by your eye care professional.

Product specifications

To learn more about our contact lenses visit:
www.coopervision.com



Product	Sphere Power	Cylinder Power (DC)	Axis°	Add Power (D)	Design	Technologies	Wear Schedule & Replacement Frequency	Base Curve (mm)	Diameter (mm)	Material USAN	Water content (%)	Oxygen transmissibility Dk/t†	UV Blocking‡ & Class	Visibility Tint	FDA Group
 Biomedics® 55 Premier	-0.25D to -10.00D (0.50D steps after -6.00D)	—	—	—	Asphere	Aberration Neutralizing System	Daily wear; 2 week replacement	8.6 8.9	14.2	ocufilcon D	55	26	Class 2	Yes	4
	+0.25D to +8.00D (0.50D steps after +5.00D)	—	—	—				8.8							
	+6.00D to -10.00D (0.50D steps after -6.00D and +5.00D)	-0.75 -1.25 -1.75 -2.25	10° to 180° (10° steps)	—				8.7							
 Biomedics® toric	+6.00D to -10.00D (0.50D steps after -6.00D and +5.00D)	-0.75 -1.25 -1.75 -2.25	10° to 180° (10° steps)	—	Toric	—	Daily wear; 2 week replacement	8.7	14.5	ocufilcon D	55	18	Class 2	Yes	4
 Proclear®	+20.00D to -20.00D (0.50D steps after +/-6.00D)	—	—	—	Sphere	PC Technology™	Daily wear; 30 days replacement	8.6	14.2	omafilcon B	62	36	No	Yes	2
	No Plano	—	—	—				8.2							
 Proclear® multifocal	+6.00D to -8.00D (0.50D steps after -6.50D)	—	—	+1.00 +1.50 +2.00 +2.50	Multifocal D lens N lens	PC Technology™	Daily wear; 30 days replacement	8.7	14.4	omafilcon B	62	14	No	Yes	2

† (@-3.00DS) x 10⁻⁹ [(cm/sec) x (ml O₂)/(ml x mm Hg)]

‡ UV-blocking contact lenses help provide protection against transmission of harmful UV radiation to eye but are not substitutes for protective UV-absorbing eyewear, such as UV-absorbing goggles or sunglasses, as they do not completely cover the eye or surrounding area. Continue to use UV-absorbing eyewear as directed by your eye care professional.

Product specifications

To learn more about our contact lenses visit:
www.coopervision.com



Product	Sphere Power	Cylinder Power (DC)	Axis°	Add Power (D)	Design	Technologies	Wear Schedule & Replacement Frequency	Base Curve (mm)	Diameter (mm)	Material USAN	Water content (%)	Oxygen transmissibility Dk/t†	UV Blocking* & Class	Visibility Tint	FDA Group
 ProcLEAR® multifocal toric	+20.00D to -20.00D (0.50D steps after +/-6.50D)	-0.75 to -5.75 (0.50D steps)	5° to 180° (5° steps)	+1.00 to +4.00 (0.50 steps)	Toric Multifocal D lens N lens	PC Technology™	Daily wear; 30 days replacement	8.7 8.4	14.4	omafilcon B	62	From 19.5 (-3.00, D, +4.00 add) To 36.8 (-3.00, D, +1.50 add)	No	Yes	2
 ProcLEAR® toric	+6.00D to -8.00D (0.50D steps after -6.50D)*	-0.75 to -2.25 (-1.25, -1.75, -2.25)	10° to 180° (10° steps)	— —	Toric	PC Technology™	Daily wear; 30 days replacement	8.8 8.4	14.4	omafilcon B	62	19.4 (MTO) 23.7 (Moulded) 16 (MTO)	No	Yes	2
 ProcLEAR® XR toric	+6.25D to +10.00D (0.50D steps after +6.50D) -8.50D to -10.00D (0.50D steps) +10.00D to -10.00D (0.50D steps after +/-6.50D)	-0.75 to -2.25 (-1.25, -1.75, -2.25)	5° to 180° (5° steps)	— —	Toric	PC Technology™	Daily wear; 30 days replacement	8.8 8.4	14.4	omafilcon B	62	13.9 10.4	No	Yes	2
 ProcLEAR® XR multifocal	+6.25D to +20.00D -8.50D to -20.00D (0.50D steps after +/-6.50D) +20.00D to -20.00D (0.50D steps after +/-6.50D) +20.00D to -20.00D (0.50D steps after +/-6.50D)	— —	— —	+1.00 to +2.50 +3.00 to +4.00 +1.00 to +4.00	Multifocal D lens N lens	PC Technology™	Daily wear; 30 days replacement	8.7 8.7 8.4	14.4	omafilcon B	62	From 19.5 (-3.00, D, +4.00 add) To 36.8 (-3.00, D, +1.50 add)	No	Yes	2

† (@-3.00DS) x 10⁻⁹ [(cm/sec) x (ml O₂)/(ml x mm Hg)]

* UV-blocking contact lenses help provide protection against transmission of harmful UV radiation to eye but are not substitutes for protective UV-absorbing eyewear, such as UV-absorbing goggles or sunglasses, as they do not completely cover the eye or surrounding area. Continue to use UV-absorbing eyewear as directed by your eye care professional.

Avaira Vitality®, Biofinity®, Biofinity Energys®, Biomedics®, clariti®, CooperVision® and ProcLEAR® are registered trademarks of The Cooper Companies, Inc. and its subsidiaries.

©2024 CooperVision 15114 06/24